



GEORGE MARLTON, JD  
PROCUREMENT DIVISION DIRECTOR

PROCUREMENT DIVISION  
PUBLIC SERVICES BUILDING  
2051 KAEN ROAD | OREGON CITY, OR 97045

REQUEST FOR PROPOSALS #2019-01  
Whole Body Digital X-Ray Scanning System- Medical Examiner's Office  
ADDENDUM NUMBER 1  
January 30, 2019

On January 15, 2019, Clackamas County ("County") published Request for Proposals #2019-01 ("RFP"). The County has found that it is in its interest to amend the RFP through the issuance of this Addendum #1. Except as expressly amended below, all other terms and conditions of the original RFP and subsequent Addenda shall remain unchanged.

**1. SECTION 3- SCOPE OF WORK; 3.3.1 Scope:**

**Remove the following product specification language:**

- The machine must be a C-arm structure, capable of rotating 0 to 90 degrees around the body to produce A/P, oblique and lateral images, without repositioning the body. It must scan in the horizontal direction so that the body does not need to be supported or secured to the table.

**Replace with the following:**

- The machine must be able to produce A/P, oblique and lateral images, without repositioning the body. In addition, the machine must scan in the horizontal direction so that the body does not need to be supported or secured to the table. The machine must have the ability to produce a complete whole body digital X-ray in a single scan.

---

End of Addendum #1